Translation of the original
operating instructions

Process gas analyser INCA4002



Union Instruments GmbH

Zeppelinstrasse 42

76185 Karlsruhe


Germany

☎ +49 (0)721-680381-0
🖷 +49 (0)721-680381-33
✉ support@union-instruments.com
💻 http://www.union-instruments.com

Item No.: 08608199955
© 2015
This documentation is copyrighted. The engendered rights are retained, in particular the rights to translation, reprinting, taking pictures, radio transmission, reproduction by photomechanical or similar methods and storage in data processing systems, including excerpts.

The right to technical changes is retained.

Dimensions


77
63
51
37
0
590
485
435
363
40
0
33,5
0
138
159,6
404,5
270
250
0
600mm
53,5mm
741mm
216mm
UNION Instruments GmbH

Measuring ranges and measuring accuracy

Refer to type plate on device also attached data and information.

Example of Measuring ranges on type plate:




Fig. 1.1: Type plate (exemplary)

1. Device description
2. Technical Information
3. Measurement ranges

Technical data

Gas inlets

Number of measuring points:	1
Calibration inlets:	1
Purge gas inlets:	1
Gas connections:	Clamp ring connection 6 mm
Max. distance between measuring point and analyser:	10 m
Max. gas inlet pressure:	20 mbar relative (optional 300 mbar)
Min. gas inlet pressure:	-100 mbar relative
Flame arrester:	ATEX certification G IIC
Relative gas humidity:	≤ 100% (condensate possible)
Condensate trap:	yes

Calibration gas

Calibration interval:	manual or automatic (configurable between one hour and up to several weeks)
Duration of calibration:	10 minutes (recommended by the manufacturer)
Gas consumption:	5 l/calibration

Power supply

Voltage:	100 - 240 V 50/60 Hz
Power consumption:	250 VA max.
Protection class:	I
Degree of protection	IP20

Interfaces

Relay:	3
Dig. interface:	RS232
Field bus:	optional
Optional relay:	optional
Remote Control Unit:	optional

Sample gas cooler

Cooling principle:	thermoelectric
Dew point:	3 - 30 °C adjustable
Condensate drain:	Hose pump

Ambient conditions

Operating temperature:	5 - 45°C
Humidity:	0 - 95 % relative humidity
Ambient pressure:	900 - 1250 hPa (0.9 - 1.2 bar)
Storage temperature:	-20 - 60°C

Weight

Weight:	up to 30 kg



⚠ ATTENTION

When using the process gas analyser in other ambient conditions, consult UNION Instruments GmbH for additional measures.

Table of contents

1 EC Declaration of Conformity 9
2 Safety notes 11
2.1 Warnings and symbols 11
2.2 Fundamentals of proper use 12
2.3 Personnel and qualifications 12
2.4 Safety notes 13
2.4.1 General notes on safety 13
2.4.2 Notes regarding special hazards 13
2.5 Regular operator training 13
2.6 Workplace hazard analysis 14
3 Safety equipment 15
3.1 Main switch 15
3.2 Safety equipment 15
3.2.1 Door - not electronically queried 15
3.2.2 Ventilator monitoring 15
3.3 Markings and warnings 16
4 Connections 17
4.1 Accessories 18
5 Transport, setup and acceptance 19
5.1 Transport 19
5.2 Ambient conditions 20
5.2.1 Storage conditions 20
5.3 Set up and connection 20
5.4 Setup site 20
5.4.1 Wall attachment 21
5.4.2 Process gas 22
5.4.3 Electrical connection 23
5.4.4 Electrical interfaces 23
5.4.5 Operator safety precautions 29
5.5 Startup after setup 29
5.6 Documentation 29
6 Startup /switching on 31
7 Description of the workplaces/operating elements 33
7.1 Workplaces 33
8 Operation 35
8.1 Description of display 36
8.1.1 Using the membrane keypad 36
8.1.2 Displayarea 37
8.2 Available displays 37
8.2.1 Menu structure 38
8.2.2 Navigate with the arrow keys left ◄ and right ► 39
8.2.3 Navigation with arrow keys up ▲ and down ▼ 40
8.2.4 Navigation with ESC and MENU 41
8.2.5 Measurement display 42
8.2.6 Measuring channel display 42
8.2.7 Saved measured values 43
8.2.8 Display in the warmup phase 43
8.2.9 Select language 44
8.2.10 Password 45

9 Decommissioning/switching off 47
10 Maintenance 49
10.1 Preparations 49
10.2 Maintenance work/Inspection 50
11 Troubleshooting 53
11.1 Preparations 53
11.2 Changing/replacing fuses 54
11.3 Messages/malfunctions on the display 54
11.3.1 Display of messages/malfunctions 54
11.3.2 Visualizing the error list 54
11.3.3 Troubleshooting list 54
12 Service 55
13 Associated documents 57
14 Disposal 59
15 Spare parts 61
16 Annex 62
Index 62
Table of figures 63

1 EC Declaration of Conformity

CE

Der Hersteller / The manufacturer

Union Instruments GmbH
Zeppelinstrasse 42
76185 Karlsruhe

erklärt hiermit, dass folgend bezeichnete Produkte / hereby declares, that following named products:

Produktbezeichnung:	Gasanalysator	Gerätegruppe:	INCA4002
Product name	Gas Analyzer	device group	INCA4002

konform sind mit den Anforderungen, die in der EG – Richtlinie festgelegt sind / are compliant with the requirements as defined in the EC directive:

2006/42/EG 2006/42/EC	Maschinenrichtlinie Machinery directive
2004/108/EG 2004/108/EC	Elektromagnetische Verträglichkeit Electromagnetic compatibility

Angewandte harmonisierte Normen / Used harmonized standards:

DIN EN 61010-1:2011	Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte - Teil 1: Allgemeine Anforderungen; Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements
DIN EN ISO 12100:2011	Sicherheit von Maschinen- Allgemeine Gestaltungsleitsätze - Risikobeurteilung und Risikominderung Safety of machinery - General principles for design - Risk assessment and risk reduction
DIN EN 61326-1:2006	Elektrische Mess-, Steuer-, Regel- und Laborgeräte - EMV-Anforderungen - Teil 1: Allgemeine Anforderungen Electrical equipment for measurement, control and laboratory use - EMC requirements -- Part 1: General requirements

Name des Dokumentationsbevollmächtigten: Name delegate of documentation	Schlichter
Adresse des Dokumentationsbevollmächtigten: address delegate of documentation	siehe Adresse des Herstellers see address of manufacturer

Bei einer nicht autorisierten Änderung des Gerätes verliert diese Erklärung ihre Gültigkeit. / Any unauthorized modification of the device results in invalidity of this declaration.

2 Safety notes

2.1 Warnings and symbols

In the operating instructions, the following names and symbols are used to denote particularly important information:

	DANGER
	Immediate danger that can lead to serious physical injury or death.

	WARNING
	Potentially hazardous situations that can lead to serious injury or death.

	ATTENTION
	Potentially hazardous situations that can lead to minor physical injury. This can also be used for property damage.

	**NOTE**
	Denotes information that can make it easier to handle the process gas analyser or help prevent property damage.

2.2 Fundamentals of proper use

The process gas analyser serves to identify gases and their quality in biogas, crude biogas, lean gas and biomethane.

Applications are biological process optimisation during motor control, controlling preparation systems, analysing biogas, landfill gas and gas from purification plants.

The gas analyser is not suitable for determining the workplace threshold or lower explosion limit.

In the case of toxic and explosive gases, observe the safety instructions at the setup site.

The process gas analyser is permanently installed and is intended for use inside closed rooms in a sufficient quantity of clean ambient air.

Any other use is considered improper. The manufacturer is not liable for the resulting damage; the associated risk is borne by the installer, fitter, operator or user. Only certified professionals may alter the process gas analyser (mechanical, electrical or pneumatic modifications).



WARNING

Proper use includes following these operating instructions. In addition to the following safety notes, always follow the safety instructions of the linked system components.

Additional equipment or accessories that are not installed, delivered or manufactured by UNION Instruments GmbH require the approval of UNION Instruments GmbH as the manufacturer! Otherwise the guarantee expires.

2.3 Personnel and qualifications

Gas connections and work on the electrical equipment of the process gas analyser may only be performed by a professional while observing safety regulations.

2.4 Safety notes

2.4.1 General notes on safety



WARNING

The process gas analyser may only be operated when all of the protective equipment is available and operable.

Additional safety notes:

☞ ***before the corresponding chapters!***

2.4.2 Notes regarding special hazards





WARNING

- **After installation, all gas conducting parts must be checked for leaks according to national regulations.**
- **All repairs that require the protective covering to be opened may only be performed by trained personnel.**
- **Sensors can contain sulphuric acid. This may leak in case of improper use. Protect from contact with skin and eyes.**

2.5 Regular operator training



NOTE

Country-specific regulations about regular user training by the operator must be observed, in particular training on handling gases and electrical equipment.

2.6 Workplace hazard analysis



NOTE

Depending on the national regulations, the operator must perform a workplace hazard analysis, if applicable independent of the CE mark for this process gas analyser.

Technical developments can give rise to deviations from these operating instructions. If you require additional information or if particular problems arise that are not fully addressed in this manual, please contact the following address:

Union Instruments GmbH

Zeppelinstrasse 42

76185 Karlsruhe

Germany

☎ +49 (0)721-680381-0
🖷 +49 (0)721-680381-33
✉ support@union-instruments.com
💻 http://www.union-instruments.com

3 Safety equipment

3.1 Main switch




Fig. 3.1: Main switch (exemplarisch)

3.2 Safety equipment

3.2.1 Door - not electronically queried

- Door of the process gas analyser.

3.2.2 Ventilator monitoring

If the housing fan fails, the process gas analyser is de-energized. The power supply unit and fan monitor control still have power.

3.3 Markings and warnings




Fig. 3.2: **Markings and warnings**

1. Type plate



Fig. 3.3: Warning in housing

4 Connections




Fig. 4.1: Product description

1. Power supply cable bushing
2. Gas cooler
3. Gas cooler fan
4. Input process gas
5. Intake air filter
6. Input calibration gas
7. Input purge gas
8. Condensate pump connection
9. Hose pump
10. Main Switch
11. Solenoid valve display, inlet on/off
12. Operating element
13. Output process gas
14. Output pressure regulator leakage
15. Cable bushings
16. Fan

4.1 Accessories



WARNING

Risk of injury/defective!

Use of non-approved accessories can cause defects and be hazardous. This will render the warranty null and void. The operator is then liable for any damage that may occur.

Only use original accessories or accessories that have been approved by Union Instruments GmbH.

5 Transport, setup and acceptance

NOTE

Generally, the process gas analyser is started up by Union Instruments GmbH or service technicians.

If it is not transported, set up and started up by Union Instruments GmbH (for example in-house transportation and resale), coordinate the appropriate procedure with Union Instruments GmbH *(☞ Chapter 12 Service)*.

5.1 Transport

⚠ WARNING





Possible injury from the process gas analyser tipping over or falling from pallets and load carrying equipment.

- **At least two persons are required to unpack and transport the analyser (for weight see technical data)!**
- **Check the load bearing capacity and condition of the slinging equipment and carefully attach it.**
- **Never stand under suspended loads.**

NOTE

In case of damage during transport from improper handling, the carrier should perform a damage report within seven days (railway, post office, freight forwarder).

5.2 Ambient conditions



⚠ ATTENTION

Ambient conditions during storage and set up.

Observe the stipulated ambient conditions. Contact Union Instruments GmbH if the process gas analyser has been stored for more than three months or needs to be operated or stored under ambient conditions other than those specified.

5.2.1 Storage conditions

If condensation water freezes in the process gas analyser, this can cause defects. Protect the process gas analyser against frost when putting it into storage.

Ambient temperature:	-20- 60°C
Humidity:	0 - 95% relative humidity
Ambient pressure:	700 - 1400 hPa (0.7 - 1.4 bar)

5.3 Set up and connection

5.4 Setup site

The setup site for the process gas analyser must satisfy the following conditions:

- Clean dry room (with the exception of INCA5000/INCA6000)
- No direct exposure to sun
- Protect from climate influences with a heater or air conditioning if necessary
- Insure a clean, sufficient amount of ambient air for undistorted measurements
- Ensure that the loadbearing capacity of the wall is sufficient



⚠ WARNING

Leaking process gas can pose a hazard and needs to be discharged by the operator into a safe environment.

5.4.1 Wall attachment

The process gas analyzer is intended for wall installation. The wall brackets are permanently attached to the housing.

The wall on which the process gas analyser is to be installed needs to be sufficiently stable to bear its weight.
Attach process gas analyser to brackets.



Fig. 5.1: Wall attachment

5.4.2 Process gas



NOTE

- **The connecting parts need to be clean and free of residue. Impurities can enter the process gas analyser and cause incorrect measurements and/or damage.**
- **The inlet pressure for the gas connections must not exceed the pressure specified on the instruction sticker on the process gas analyser.**
- **Each connection needs to be carefully checked for leaks. If there are any leaks, the system will draw air, and the measurements will be incorrect.**
- **Do not use sealing compound to seal the gas connections as this can lead to inaccurate measurements. Use PTFE sealing tape.**
- **Only use suitable pipes.**
- **Use a separate line to drain off the condensate.**



⚠ ATTENTION

The process gas must be free of condensate and dust if the process gas analyser has no gas preparation system (or gas cooler).

5.4.3 Electrical connection



DANGER

Danger from electrical shock!

Only a trained electrician may modify the electrical equipment of the process gas analyser in accordance with the relevant guidelines.

When the process gas analyser has been opened, the parts identified by the adjacent symbol may still be live even when the master switch has been turned off. If necessary, disconnect the process gas analyser from the power supply.

5.4.4 Electrical interfaces



WARNING

Untrained personnel starting the process gas analyser may endanger people and equipment.

Only trained service technicians may start up the analyser.



NOTE

Only operate relay with functional extra low voltage.

Do not connect to the power supply.

Power Supply Connections




Fig. 5.2: Power Supply X2 with fuses

1. Power supply connections

Connect the process gas analyser to the power supply in accordance with national regulations via L1, N and PE.

Electrical Interfaces




Fig. 5.3: Electrical interfaces

Item No.	Designation
1	Relay X10A☞ *Fig. 5.4* and *5.5!*
2	Analogue outputs X11A (optional) ☞ *5.6*
3	Profibus module X12 (optional)

Relay





Fig. 5.4: Relay X10A, outputs K1 – K3

Item No.	Designation	Function
1	Relay K1	Operation
2	Relay K2	Failure (inverted)
3	Relay K3	OFF



K_n-NC – K_n-COM – K_n-NO

Fig. 5.5: Relay terminal assignment

left - middle: normally closed
right - middle: normally open



NOTE

Only operate relay with functional extra low voltage.

Do not connect to the power supply.

Maximum load of the relay connections 30VDC / 1A.

Analogue outputs




Fig. 5.6: Analogue outputs X11A, connections

Item No.	Function	Item No.	Function (2-channel device)
1	Output 1 – signal/signal 4-20 mA	**5**	Output 5 – signal/signal 4-20 mA
-	1 GND	**-**	5 GND
2	Output 2 – signal/signal 4-20 mA	**6**	Output 6 – signal/signal 4-20 mA
-	2 GND	**-**	6 GND
3	Output 3 – signal/signal 4-20 mA	**7**	Output 7 – signal/signal 4-20 mA
-	3 GND	**-**	7 GND
4	Output 4 – signal/signal 4-20 mA	**8**	Output 8 – signal/signal 4-20 mA
-	4 GND	**-**	8 GND

Assignment of analog interface configurable with Software INCACtrl.

The load resistor is 500 ohm.

Remote Control Unit RCM




Fig. 5.7: Remote Control Unit RCM X15

Pos.-No.	Description
1	bus, connecting internal power supply
2	internal, RS232 connection for PCB-AddOn (Display) via null modem cable
3	connection Fieldbus coupler
4	bus, connecting internal power supply
5	DIP switches
6	Ethernet
7	status LED, LED 1- USB active, LED 2- Fieldbus active, LED 3 - Ethernet active (from top to bottom)
8	Micro-USB, local connection to PC, if used no connection via Ethernet/Fieldbus

Communication module for integration into Ethernet networks to communicate and operate the process gas analyser.

DHCP is enabled as factory setting, RCM received IP - address automatically from a DHCP server. Manual assignment of IP address with separate software (example: "DeviceInstaller", Lantronix).

Default Settings:

IP over DHCP	On
Port	10001
Protocol	TCP/IP
RS232	115200 bit/s, 8 data bit, 1 stop bit, no parity bit

MAC address of RCM, refer to label on RCM.

5.4.5 Operator safety precautions



WARNING

- **The operator needs to provide suitable safety equipment for the process gas analyser to reliably prevent individuals from being injured from gas leaks.**
- **Any leaking process gas needs to be diverted into a safe environment.**
- **Identify the exit point of the diverted gas with a warning.**
- **Danger of tripping due to improperly installed supply lines.**

Install the supply lines in a suitable manner.

5.5 Startup after setup



WARNING

Untrained personnel starting the process gas analyser may endanger people and equipment.

Only trained service technicians may start up the analyser.

5.6 Documentation



NOTE

Union Instruments GmbH recommends keeping a maintenance manual and documenting all jobs and tests.

6 Startup /switching on



⚠ ATTENTION

To establish operational readiness, including of the linked system components, according to the corresponding operating instructions.



NOTE

The following table contains highly abbreviated steps for starting up after a long downtime.

To turn on the process gas analyser after a short downtime, a few steps can be omitted:
☞ ***right column!***

Steps	Startup	Turning on
Check whether the ambient conditions (☞ *Technical Data chapter on page 5*!) meet the requirements.	X	X
Check that the process gas analyser has been fastened securely.	X	
Check that the device is suitable for the process gas.	X	
Check that the process gas is correct.	X	
Check that the gas connections are correct and tight.	X	X
Check the integrated filters (water/fine filter) for condensate, if necessary.	X	X
Check, if necessary, that the calibration gas is correct.	X	X
Establish/switch on the operator energy and media supply.	X	X
Check the voltage.	X	
Open shut-off valves.	X	X
Turn on the master switch.	X	X
Make sure the linked system components are ready to start.	X	X
If the process gas analyser was only switched off temporarily, production can be resumed.		

7 Description of the workplaces/operating elements

NOTE

This chapter only discusses the elements used by normal operators to operate the process gas analyser.

7.1 Workplaces



Fig. 7.1: Workplaces

Item No.	Designation	Function/Activity
1	Display with status LED	Display status.

Display status LED

The following states are displayed through those LEDs:

LED Operation	
Output state	**Description**
flashing	Device functionality OK (even Service might be pending)
flashing	Device functionality is affected by errors, Service message pending
flashing	Device stopped by fatal error, Error pending

Fig. 7.2: Status LED

8 Operation



WARNING

Danger of injury!

Only use the process gas analyser when all lines have been installed and checked for leaks according to national regulations.

8.1 Description of display

8.1.1 Using the membrane keypad

The software controls are operated using a membrane keypad. The displayed buttons can be selected by pressing the key. The menu structures are intentionally flat to enable quick access to functions.

⚠ ATTENTION

Damage to the membrane keypad!

The membrane keypad may be damaged if you use other objects to operate it apart from your fingers.




Fig. 8.1: Operating elements

Item No.	Designation	Function
1	Measurement display	Display the current sensor measurements.
2	Measuring channel display	Display the current channel measurements.
3	Saved measured values	Switch between the last 10 saved measured values.
4	Display	Display values, times and measurement results
5	Menu keys	Navigating the menu structure
6	Status LED	Display state of device

8.1.2 Displayarea




Fig. 8.2: Display area

Item No.	Designation	Function
1	Top display area	Display the status and channel information
2	Bottom display area	Switch between various measured values with the arrow keys (▼▲►◄).

8.2 Available displays

NOTE

The available displays and corresponding functions are described below. The navigation path to the displays is indicated by the menu and function keys in the chapter headings.

The control system is based on the structure shown below.

8.2.1 Menu structure



NOTE


If some of the menu items (framed in red) are changed, this can subsequently influence the measurement results.

Main menu
- **Settings**
  - Language
  - Password
  - Output data
  - Screen change
- **Parameter**
  - ABC built-in
  - EC meas. Cycle[1)]
  - Purge time[1)]
- **Commands**
  - Start measurement
  - Stop measurement
  - Restart System
  - Clear messages
  - Calib. purge gas
  - Calib. gas 1
  - Calib. gas 2[1)]
  - Reset cal.-data
  - Test (gas 1)
  - Abort calib.
  - Check OK
- **System info**
  - Version firmware
- **System messages**


The menu structure refers to firmware version V1.08.


[1)] only available for certain device configurations

8.2.2 Navigate with the arrow keys left ◄ and right ►



- The display indicates that the measuring status is active.





- Switch between the measurements by pressing the left ◄ and right ► keys.



The asterisk (*) indicates that a saved value is being displayed. The values are updated in the display depending on the measuring status.


With continuous measurement, an asterisk is not displayed since the value is measured and updated continuously.

8.2.3 Navigation with arrow keys up ▲ and down ▼

NOTE

To navigate with the arrow keys up ▲ or down ▼, use the left ◄ and right ► arrow keys to select the display in which the date and time are shown.

```
MEASURING            4:57
Channel   :           1/3
        07/22/2009
         14:42:21
```



- Press the up ▲ and down ▼ keys to display other data.

```
MEASURING            4:53
Channel   :           1/3
Err       :       0 Errors
MSGS      :     7 Messages
```

- "Err" displays the number of saved errors.
- "MSGS" shows the number of saved messages.

```
MEASURING            4:50
Channel   :           1/3
pAir      :      1.8 mbars
pGas      :      0.3 mbars
```

- "pLuft" and "pGas" are the differential pressures measured for the individual gas pathways (air and process gas) in the process gas analyser.

```
MEASURING            4:45
Channel   :           1/3
T_IR      :      49.2  °C
TCool     :       5.3  °C
```

- "T_IR" is the current temperature of the infrared measuring unit.
- "TCool" is the current temperature of the gas cooler.

```
MEASURING            4:45
Channel   :           1/3
TCase     :      49.2  °C
Tout      :       5.3  °C
```


- "TCase" is the current temperature in the housing.
- "Tout" is the current ambient temperature.

8.2.4 Navigation with ESC and MENU






- With the MENU key to the main menu.






- Select the submenu with the up ▲ and down ▼ keys.
- Confirm the selection by pressing the MENU key.



- Press the ESC key in the menu to go one level higher.

8.2.5 Measurement display







- Select the individual measurements with the plus **+** or minus **-** keys.

The asterisk (*) indicates that a saved value is being displayed. The values are updated in the display depending on the measuring status.

With continuous measurement, an asterisk is not displayed since the value is measured and updated continuously.


8.2.6 Measuring channel display





- Use the up ▲ or down ▼ keys to select the individual channels.


8.2.7 Saved measured values





- Press the forward ↻ or back ↺ key to step through the last saved measured values.
- The measured values are identified by the count/date/time.

8.2.8 Display in the warmup phase




The figure shows the display during the warm-up phase. In the figure, the infrared electronics, T(IR), have reached operating temperature. Depending on the type of sensor, this is 49°C or 64°C. The Parox sensor, T(POX), is not ready. Once it reaches its operating temperature, the display shows T(POX)=0x0000 and OK.


Devices with sensors that do not require a specific operating temperature start without a warm-up phase and start measuring immediately when switched on.

8.2.9 Select language





- Select the language with the ▼▲ keys.
- Confirm the selection by pressing the MENU key.





Available languages: German, English, Italian and Spanish




8.2.10 Password

⚠ ATTENTION

The password has a maximum of four characters.

If you forget the password, you cannot change the configuration.







- Enter the password using the▼▲►◄ keys.
- Confirm the entry by pressing the MENU key.



9 Decommissioning/switching off



⚠ ATTENTION

To decommission the process gas analyser and the linked system components according to their operating instructions.



NOTE

The following table contains the steps for decommissioning the analyser for a long period.

If you only wish to switch off the process gas analyser temporarily, a few steps can be omitted:
☞ ***Switch off column!***

Steps	Turn off	Decommissioning
Disconnect the device from the process, close the line professionally.	X	X
Rinse the process gas analyser with ambient air. (Start calibration with purge gas)		X
Shut down the linked system components.	X	X
Turn of the master switch.	X	X
 If you only wish to switch off the process gas analyser temporarily, follow the procedure here to the end!		
If required, disconnect / switch off the operator's energy and media supply and the signal transmission professionally.		X
If advantageous, pack process gas analyser.		X

10 Maintenance

The measuring quality of the process gas analyser can only be ensured if the service intervals are maintained.

10.1 Preparations


The feed lines to linked system components can be closed for servicing purposes. Once operation has been resumed, they need to be reopened.



DANGER

Serious risk of injury from electricity.

- **Parts of the process gas analyser labelled with this symbol may still be live even when the main switch has been switched off. If necessary, disconnect the process gas analyser from the power supply.**
- **Switch off the main fuse and if necessary, secure it to prevent it from switched back on.**
- **Only a trained electrician may work on the electrical equipment of the process gas analyser.**



WARNING

Serious risk of injury from exiting gas.

- **Switch off the process gas analyser, and also linked system components if required, before carrying out maintenance work.**
- **The gas connections may only be established by trained personnel. Follow the applicable guidelines at the installation site.**

10.2 Maintenance work/Inspection



NOTE

The maintenance work must be carried out in accordance with the inspection and maintenance schedule! The type and extent of the wear depends on the individual application and operating conditions. All intervals specified are therefore for guidance only.

Upon reaching the inspection intervals, this will be shown in the display. Conduct and document testing. Through menu confirm that validation was done: MENU→COMMANDS→Check OK→ [Enter]

Inspection	Interval (recommended)	
Weekly inspection		
Purge gas inlet unobstructed (particularly in case of frost)	weekly	
Exhaust gas line unobstructed (particularly in case of frost)	weekly	
Quarterly inspection		
Calibrate device according to manufacturer's specifications, message in display “Service [Typ] ZERO”, “Service [Typ] SPAN”, “Service [Typ] MID”	every 3 months, latest annually or when required	
Half-yearly inspection		
Check integrated filter in the device	every 6 months	
Check compressed air supply[1)] (negative pressure during "drainage")	every 6 months	
Check lines for condensate (including all integrated filters)	every 6 months	
Check gas inlets and clean professionally if necessary	every 6 months	
Check fan	every 6 months	
Check ejector pump[1)]	every 6 months	
Check inlet filter (inlet ambient air, filter mat, ventilator)	every 6 months	
Check Peltier cooler[1)]	every 6 months	
Check fan of Peltier cooler[1)]	every 6 months	
Yearly inspection		
Check air and gas pump[1)] (by performing a purge gas calibration)	yearly	

[1)] if installed

Maintenance/Replacing components	Interval (recommended)	
Half-yearly service and after commissioning		
Check and, if necessary, update firmware version	every 6 months	
Save the current configuration with INCACtrl	every 6 months	
Annual service		
Replace integrated filters	every 12 months	
2-yearly service		
Replace pump hoses	every 24 months	
Replace flame arrester[1)]	every 24 months	
8-yearly service		
Replace integrated pressure reducer	every 8 years	
If necessary		
Replace gas-delivering pumps	if necessary	
Replace sensor, lifetime depends on sensor type, message in display “Service [Typ] age” or “Service [Typ] usage”	if necessary	

[1)] if installed

11 Troubleshooting



NOTE

A distinction is made between the following categories:

Possibility of the measured values:

Measured values that deviate from the anticipated range

☞ *last maintenance!*

☞ *maintenance manual!*

Faults:

Faults during operating process

For remedy: ☞ *Section 11.2, p.* 54

11.1 Preparations

The feed lines to linked system components can be closed for servicing purposes. Once operation has been resumed, they need to be reopened.







⚠ WARNING

Serious risk of injury from electricity and exiting gas.

- **Switch off the process gas analyser, and also linked system components if required, before carrying out maintenance work.**
- **Switch off the main fuse and if necessary, secure it to prevent it from switched back on.**
- **Only a trained electrician may work on the electrical equipment of the process gas analyser.**
- **Parts of the process gas analyser labelled with this symbol may still be live even when the main switch has been switched off.**
- **If necessary, disconnect the process gas analyser from the power supply.**

11.2 Changing/replacing fuses

Fuses may only be exchanged by an electrician or service professional. Choose the type approved by Union.

11.3 Messages/malfunctions on the display

11.3.1 Display of messages/malfunctions

If errors occur during operation, the control system automatically switches to overview to display priority messages.

11.3.2 Visualizing the error list

NOTE

The following list contains errors and messages that are visually displayed.

Troubleshooting measures:

☞ ***Troubleshooting list!***

Measures required are assigned via column [☞ *no.*].

Error text	Error message	☞*No.*
0x30D	Minimum pump pressure not reached, Sensor EC, Pressure Air	1
0x30E	Minimum pump pressure not reached, Sensor EC, Pressure Gas	2
Additional	All additional	3

11.3.3 Troubleshooting list

The following list contains causes of faults.

No.	Description
1	Inlet air filter for ambient air clogged (Fig. 4.1)
2	Process gas outlet blocked, for example frozen (Fig. 4.1) Process gas inlet closed (Fig. 4.1) - too much condensate in the line
3	Contact service ☞ *Chapter 12!*

12 Service



NOTE

If you have any questions Union Instruments GmbH will be happy to assist.

In case of orders or technical questions, please have the customer number, telephone number for return calls, the type and number of the process gas analyser (see the type plate) and the required spare parts and parts list numbers to hand.

Union Instruments GmbH - Service

Maria-Goeppert-Straße 22

23562 Lübeck

Germany

☎ +49 (0)721-680381-30
✉ support@union-instruments.com
💻 http://www.union-instruments.com

13 Associated documents

- Declaration of Conformity for flame arrester [1)]
- Operating and service log
- Service documentation, optional

[1)] if installed

14 Disposal

Following decommissioning, the analyser can be returned to Union Instruments GmbH.

Suggestion, have Union Instruments GmbH dispose of the process gas analyser.



WARNING

Serious risk of injury from electricity and exiting gas in the process gas analyser.

Before disassembly, disconnect the process gas analyser from the energy supply.

If necessary, purge the gases.



NOTE

Observe the national regulations on disposing of machines and operating materials!

Grade the parts according to group and recycle properly.

15 Spare parts



WARNING

The use of non-approved spare parts (such as parts from other manufacturers, parts with different specifications, replicas of used and wear parts) can cause defects and be hazardous. This will render the warranty null and void. The operator is liable for any damage that occurs as a result.

When replacing standard components, only use identical components by the original manufacturer. If components are discontinued or components by different manufacturers are used, request the manufacturer approval by Union Instruments GmbH.

Spare parts can be ordered from Union Instruments GmbH:
☞ *Chapter 12 Service*.

Note the type and number *(☞ type plate)* of the process gas analyser.

If necessary, find and make a note of the order number *(☞ Applicable documents).*

Order part.

16 Annex

Index

A
Accessories 18
Ambient conditions..... 20
C
Connecting the process gas analyser ... 20
Connections..... 17
Contact
  Service 55
  Union Instruments GmbH 14
D
Decommissioning.....47, 59
Display..... 37
Displays 37
Disposal..... 59
E
EC Declaration of Conformity..... 9
Electrical connection 23
Electrical interfaces..... 23
Error elimination..... 53
I
Inspection 50
L
load resistor 27
M
Main Switch 15
Maintenance 49
Maintenance work..... 50
Measurement display..... 42
Measuring channel display..... 42
N
Navigation with arrow keys 39, 40
Navigation with ESC and MENU.....41
O
Operating elements 33
Operation 35
  Membrane keypad 36
  with arrow keys 39, 40
  with ESC and MENU.....41
P
Password 45
Personnel and qualifications.....12
Process gas.....22
Proper use.....12
S
Safety equipment 15
Safety notes 11, 13
Saved measured values 43
Select language.....44
Service 55
Setting up the process gas analyser.....20
Setup site 20
Spare parts.....61
Startup.....31
Symbols 11
T
Transport.....19
Troubleshooting.....53
W
Wall attachment.....21
Warm-up phase.....43
Warnings 11
Workplaces 33

Table of figures

Fig. 1.1: Type plate (exemplary) 4
Fig. 3.1: Main switch (exemplarisch) 15
Fig. 3.2: Markings and warnings 16
Fig. 3.3: Warning in housing 16
Fig. 4.1: Product description 17
Fig. 5.1: Wall attachment 21
Fig. 5.2: Power Supply X2 with fuses 24
Fig. 5.3: Electrical interfaces 25
Fig. 5.4: Relay X10A, outputs K1 – K3 26
Fig. 5.5: Relay terminal assignment 26
Fig. 5.6: Analogue outputs X11A, connections 27
Fig. 5.7: Remote Control Unit RCM X15 28
Fig. 7.1: Workplaces 33
Fig. 7.2: Status LED 34
Fig. 8.1: Operating elements 36
Fig. 8.2: Display area 37